



Dependable Products From People You Trust

11302 Steeplecrest Drive ~ Houston, Texas 77065 ~ Phone (832) 320-7300 ~ Fax (713) 880-9886 ~ Web www.ofite.com

## OFI TESTING EQUIPMENT, INC. CERTIFICATE OF CONFORMANCE

This is to certify that the following instrument has been manufactured and validated per the specifications of OFI Testing Equipment, Inc., and meets or exceeds the following requirements.

*API RP 13B-1, Recommended Practice for Field Testing of Water-based Drilling Fluids*  
§ 7.3.1.2.1 High-temperature/high-pressure (HTHP) test – Apparatus – Filter paper

*API RP 13B-2, Recommended Practice for Field Testing of Oil-based Drilling Fluids*  
§ 8.2.1.2 a) High-temperature/high-pressure (HTHP) test up to 175°C (350°F) –  
Apparatus – Filter paper

Our Filter Paper, Part Number 170-19, is manufactured by our supplier using Whatman Grade No. 50 Filtrate Paper.

PART NO. 170-19

DESCRIPTION Filter Paper, API, Whatman Grade #50 2 1/2"

QUALITY ASSURANCE \_\_\_\_\_



**КОПІЯ ВІРНА**

ДИРЕКТОР  
ГОРБАЧОВ Ю. А.



*Dependable Products From People You Trust*

11302 Steeplecrest Drive ~ Houston, Texas 77065 ~ Phone (832) 320-7300 ~ Fax (713) 880-9886 ~  
Web [www.ofite.com](http://www.ofite.com)

## OFI TESTING EQUIPMENT, INC.

### СЕРТИФІКАТ ВІДПОВІДНОСТІ

Цим засвідчується, що наступний прилад виготовлений та перевірений відповідно до специфікацій компанії OFI Testing Equipment, Inc. та відповідає наступним вимогам або перевищує їх.

*API RP 13B-1, Рекомендована практика польових випробувань бурових рідин на водній основі*

§ 7.3.1.2.1 Випробування на високу температуру / високий тиск (НТНР) - Апарат - Фільтрувальний папір

*API RP 13B-2, Рекомендована практика польових випробувань бурових рідин на масляній основі*


§ 8.2.1.2 а) Випробування на високу температуру / високий тиск (НТНР) до 175 ° C (350 ° F) - Апарат - Фільтрувальний папір

Наш фільтрувальний папір, номер за каталогом 170-19, виготовляється нашим постачальником з використанням фільтрувального паперу Ватман марки 50.

ПАРТІЯ № 170-19

ОПИС Фільтрувальний папір, API, Ватман марки #50 2 1/2"

Контроль якості \_\_\_\_\_



ДИРЕКТОР  
ГОРБАЧОВ Ю. А.



Переклад виконано з англійська мови  
на українське мову  
Прозито і пронумеровано 2 (962) аркушів  
Відповідальна особа:  
перекладач: /П/ - Горбачова О.Г.  
Ім'я Прізвище